



Babycare TENS®

Accompanying Electromagnetic Information

This product needs special precautions regarding EMC and needs to be installed and put into service according to the EMC information provided. The unit can be affected by portable and mobile RF communications equipment.



Caution: This unit has been thoroughly tested and inspected to assure proper performance and operation!

Caution: This device should not be used adjacent to or stacked with other equipment. If adjacent or stacked use is necessary, the device should be observed to verify that it is operating normally.

Body Clock Health Care Ltd

108 George Lane, South Woodford, London
E18 1AD United Kingdom

Tel (+44) (0)20 8532 9595 Fax (+44) (0)20 8532 9551

www.bodyclock.co.uk

e-mail customercare@bodyclock.co.uk



This leaflet is recyclable

July 2016 (2)

Guidance and manufacturer's declaration - electromagnetic emissions

This unit is intended for use in the electromagnetic environment specified below.
The customer or the user of this unit should ensure that it is used in such an environment.

Emissions test	Test	Electromagnetic environment – guidance
RF emissions CISPR 11	Group 1	This unit uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference with nearby electronic equipment.
RF emissions CISPR 11	Class B	
Harmonic fluctuations/ IEC 61000-3-2	n/a	This unit is suitable for use in all establishments other than those directly connected to the public low-voltage power supply networks that supplies buildings used for domestic purposes.
Voltage fluctuations/ flicker emissions IEC 61000-3-3	n/a	

Recommended separation distances between portable and mobile RF communications equipment and this unit

This unit is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of this unit can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and this unit as recommended below, according to the maximum output power of the communications equipment.


Rated maximum output power of transmitter (W)	Separation distance according to frequency of transmitter (m)		
	150kHz to 80MHz $d=1.2xP^{1/2}$	80MHz to 800MHz $d=1.2xP^{1/2}$	800MHz to 2.5GHz $d=2.3xP^{1/2}$
0.01	0.12	0.12	0.23
0.1	0.38	0.38	0.73
1	1.2	1.2	2.3
10	3.8	3.8	7.3
100	12	12	23

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in metres (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

- i** Note 1: At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.
- i** Note 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

Guidance and manufacturer's declaration - electromagnetic immunity

This unit is intended for use in the electromagnetic environment specified below. The customer or the user of this unit should ensure that it is used in such an environment.

Immunity test	IEC 60601	Compliance	Electromagnetic environment guidance
Conducted RF IEC 61000-4-6	3 Vrms 150kHz to 80MHz	n/a	Portable and mobile RF communications equipment should be used no closer to any part of this unit including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended Separation Distance $d = [3.5 / \sqrt{P}] \times P^{1/2}$ $d = 1.2 \times P^{1/2}$ 80MHz to 800MHz $d = 2.3 \times P^{1/2}$ 80MHz to 2.5GHz
Radiated RF IEC 61000-4-3	3V/m 80MHz to 2.5GHz	3V/m	Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey should be less than the compliance level in each frequency range. Interference may occur in the vicinity of equipment marked with the following symbol: 

① Note 1: At 80MHz and 800MHz, the higher frequency applies.

① Note 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

- a). Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which this unit is used exceeds the applicable RF compliance level above, this unit should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating this unit.

b). Over the frequency range 150 kHz to 80 MHz, field strengths should be less than $[V_1] V/m$.

Guidance and manufacturer's declaration - electromagnetic immunity

This unit is intended for use in the electromagnetic environment specified below. The customer or the user of this unit should ensure that it is used in such an environment.

Immunity test	IEC 60601	Compliance	Electromagnetic environment guidance
Electrostatic discharge (ESD) IEC 61000-4-2	±6kV contact ±8kV	±6kV contact ±8kV	Floor should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transient / burst IEC 61000-4-4	±2kV for power supply lines ±1kV line and neutral	n/a	n/a
Surge IEC 61000-4-5	±2kV to line(s) to line ±1kV line(s) to earth	n/a	n/a
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	<5% UT (95% dip in UT) for 0.5 cycle 40% UT (60% dip in UT for 5 cycles) 70% UT (30% dip in UT) for 25 cycles <5% UT (95% dip in UT) for 5s	n/a	n/a
Power frequency (50/60Hz) magnetic field IEC 61000-4-8	3A/m		Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.

i NOTE UT is the a.c. mains voltage prior to application of the test level